











Abutment/šroub	Točivé momenty in Ncm				
	10	15	20	25	30
Vhojovací kapna Tvarovač gingivy Tvarovač gingivy, úzký 	PF 3.0				
	PF 3.5				
	PF 4.0				
	PF 4.5				
	PF 5.0				
	PF 6.0				
Šrouby abutmentu 		PF 3.0			
		PF 3.5			
				PF 4.0	
				PF 4.5	
				PF 5.0	
Dynamický šroub abutmentu 		PF 3.5			
				PF 4.0	
				PF 4.5	
				PF 5.0	
Šroub abutmentu v kombinaci s abutmentem ART 		PF 3.5			
			PF 4.0		
			PF 4.5		
			PF 5.0		
Abutment VARIOmulti 		PF 3.5			
				PF 4.0	
				PF 4.5	
				PF 5.0	
				PF 6.0	
Ochranná kapna VARIOmulti 	PF 3.5				
	PF 4.0				
	PF 4.5				
	PF 5.0				
	PF 6.0				
Okluzní šroub (VARIOmulti, VARIO 17°, třmen) Uzavírací šroub pro třmen CAD/CAM 		PF 3.5			
		PF 4.0			
		PF 4.5			
		PF 5.0			
Kulová kotva 				PF 4.0	
				PF 4.5	
				PF 5.0	
			PF 3.5		
Abutment ZEST LOCATOR® 					PF 4.0
					PF 4.5
					PF 5.0
					PF 6.0
Abutment ZEST LOCATOR® a Límec pro VARIOmulti 			PF 4.0		
			PF 4.5		
			PF 5.0		
			PF 6.0		

Pro definitivní začlenění suprakonstrukce do ústí je pro fixování třeba použít vždy nové šroubky abutmentů.